

INTELLIGENT PHONOCARDIOGRAM ANALYSIS AND REPRESENTATION TOOL

J. P. Ramos, P. Carvalho, R. P. Paiva, L. Vale, J. Henriques

Department of Informatics Engineering, University of Coimbra, Pólo II, Coimbra, Portugal
jpramos@student.dei.uc.pt, {carvalho, ruipedro, lluis, jh}@dei.uc.pt

Keywords: Phonocardiogram Analysis, eHealth, Intelligent Stethoscope.

Abstract: Cardiac auscultation is a highly sensitive, specific, cost effective and comfortable diagnosis technique for many cardiovascular diseases. Unfortunately, it is observed that the art of auscultation is mastered by an increasingly lower number of medical professionals. This paper presents a Matlab tool to support physicians in performing auscultation. This application enables real time signal acquisition using off-the-shelf sensors and performs several automatic annotation functions of heart sounds, such as noise contamination detection, segmentation into S1, S2 and S3, S2-split detection, murmur detection and classification, systolic time intervals measurement, contractility and stroke volume. These are related to the most pertinent clinical applications of this signal. Moreover, it can also be used for auscultation training.

1 INTRODUCTION

Cardiovascular diseases (CVD) are a major health problem in Europe, causing 42% of all deaths in the European Union (EU). On top of that, CVD is estimated to cost the EU economy €192 billion a year. Moreover, 54% of that cost can be accounted for by the expenses of an inpatient hospital care for people who have CVD and another 28% to drugs (Allender et al., 2008).

To prevent such high costs in health care, a measure that can be pursued, aside from fostering a healthy lifestyle, is to prevent and control CVDs at an early stage. Moreover, more than just use any diagnostic tool available, such as an echocardiography, which leads often to unnecessary and inefficient use of resources, one ought to seek sources of diagnosis that can equally make an accurate referral decision and, at the same time, be less expensive (Shub, 2003).

The heart sounds have been carefully studied and related to physiological events within the heart (Wattous, 2006). Either by the presence of a specific heart sound or its acoustic properties, one is able to infer important diagnostic analysis. Characteristics such as timing, relative intensity and frequency, form the basis of auscultatory and phonocardiographic diagnosis

of CVDs.

However, cardiac auscultation has its disadvantages. First and foremost, by having their dominant frequencies below the threshold of hearing, the heart sounds are barely audible; secondly, auscultation is dependent on the physicians' judgement, therefore the diagnosis is subjective; last but not least, studies have shown that medical schools have been disregarding the teaching of auscultation (Mangione et al., 1993), and that cardiac examination skills may decline over the years mostly due to a lack of continuous training (Lam et al., 2005; Vukanovic-Criley et al., 2006).

To overcome these difficulties, computer-aided auscultation attempts to assist a general practitioner in judging an appropriate referral. By means of signal processing techniques, algorithms have been developed for elementary processing function, such as heart sound segmentation and murmur classification, and more recently for systolic time measurements, i.e. the pre-ejection period (PEP) and the left ventricle ejection time (LVET), and cardiac function indexes such as cardiac output, stroke volume and contractility. Although several frameworks can be found in the literature that tackle the former (Rajan et al., 1998; Kudriavtsev et al., 2007), none of the aforementioned tackle the latter processing functions.

Some applications can be found in the literature. In (Reed et al., 2009) a software application is presented that displays graphical representations of heart sound signals and manage existing acquisitions, but still lacks the ability to make acquisitions - recordings are uploaded to the application - and identify heart sound components or other information from phonocardiograms (PCGs), as this is currently being made by cardiologists that listen to and describe the sounds. A commercially available service is briefly described in (Watrous, 2006). Although this has more features than the previous one, the processing module only provides information regarding segmentation results and murmur identification, lacking information such as systolic time measurements and cardiac functions assesment, which also provide rich information to a more accurate decision by the physician.

In this paper we introduce an intelligent stethoscope (fig. 2) implemented with an off the shelf sensor and a Matlab tool for the acquisition, analysis, management and visualisation of cardiac signals. We will mainly focus in the last three features of the application (see figure 1), whereas a more detailed description of the processing module can be found in (Carvalho et al., 2011). The goals behind this tool are to provide not only detailed information on the heart's function to a physician in his/her diagnostic referral, but also a learning tool or a skill trainer to any medical trainee.

The paper is organised as follows: in Section 2, the signal acquisition toolbox is described. Section 3 presents the data management module and section 4 delineates the user interface layer of the application. Finally, conclusions and future work are discussed in section 5.




Figure 1: Layer Architecture of the application.

2 SIGNAL ACQUISITION TOOLBOX

The cardiac signal acquisition process plays an important role in this solution. It might not only be necessary to acquire heart-sound signals but also ECG signals in order to feed the processing layers with the

necessary data to assess a patient cardiac condition. Furthermore, the need to integrate off-the-shelf or already available cardiac signal sensors is of absolute importance, since the platform adoption on behalf of the physicians depends heavily on its ability to make use of already available equipment.

2.1 Sensor Middleware

Given the signal acquisition needs, we decided to adopt an already available sensor middleware solution (Brito et al., 2010) that was developed inside our research group.

The main features supplied by this component are: (i) off-the-shelf heterogeneous sensor integration; (ii) ease of deployment, since we only have a single software component; (iii) data relaying capabilities between middleware instances; (iv) service access transparency; (v) integration of data processing routines, such as diagnosis support algorithms; (vi) sensor discovery, registry and admission features as well as communication Quality of Service (QoS) capabilities, such as management of communication channels persistence.

Depicted on figure 3 we can see a simplified diagram of the Signal acquisition toolbox architecture. This includes the native sensor middleware as well as a Matlab API.

Under the native sensor middleware, the integration of an arbitrary number of off-the-shelf devices is accomplished with the definition of Data-Source and Data-Sink abstractions in the various layers of the middleware. The data provided by the sensors is made available to the upper layers through the use of a service-oriented approach where a consumer can subscribe and unsubscribe any given number of services published in the middleware component. The service subscription process requires that the consumer provides a Data-Sink to where the data will be relayed as soon as they become available. Another functionality relies on the ability to associate data processing schemes to available services and providing them seamlessly to the consumer. In this way, one can access non-existent raw sensor data through the processing of existing services, i.e. given the unavailability of ECG quality evaluation mechanisms or Heart Rate (HR) estimation and the existence of an ECG service it is possible to determine this missing parameter and publish it as a service for consumption.

Data logging capabilities are also present. A consumer can specify the services that should be stored. Access to the stored data can be made by subscribing to the services published by the data logger on the middleware service layer.




Figure 2: Main Window of the application where the user can find (i) patient information, (ii) relevant auscultation information and (iii) graphical representations of cardiac signals.

Also, given the component modularity it is possible to expand its communication capabilities since the middleware adopts communication protocol plug-in mechanisms. In this way, this component presents itself as a highly expandable and flexible solution allowing for the integration of an arbitrary number of communication protocols. As of now, it already provides support for TCP/IP, UDP/IP, Bluetooth and 802.15.4 with several data exchange protocols, but it might also integrate the IEEE 11073 protocol in order to support medical device systems, enabling the middleware to act as a distributed signal aggregator in a x73 network.

One of the drivers implemented was a generic audio driver that acts as a Data-Source. This driver is perfectly suited to acquire Heart-Sound signals from available audio devices and is able to handle multi-channel acquisition of audio up to 44100Hz.

2.2 Matlab Interface Solution

Having addressed the signal acquisition needs with the adopted sensor middleware, which was developed in C++ (and since the tool depicted in this paper focuses a Matlab environment), the integration of the acquisition component was required. For this task several integration techniques supported by Matlab were considered.

First, the definition of a Java API that could be accessed directly by Matlab on one side and a JNI in-

terface to access the middleware native code on the other was a possibility. However this would introduce some overhead since the data would have to be translated from C++ to Java and from Java to Matlab. Another problem with this approach relies on the fact that Matlab requires the use of a specific version of Java Virtual Machine (JVM) so future compatibility could be an issue.

Another approach focused on the use of Matlab Executables (MEX). This technique allows that dynamically linked subroutines produced from C, C++ and Fortran can be called from within Matlab. This brings evident advantages compared with the Java approach since we will only need to define one API simplifying the implementation process but also reducing the data translation overhead. Moreover, in terms of future compatibility as well as retro-compatibility we would no longer need to worry about third-party software since the MEX integration technique was created by Mathworks itself and has had a long-lasting support throughout Matlab releases.

So, as for this integration task, the MEX approach was implemented.

As we stated above, the MEX approach requires the definition of dynamically linked subroutines (DLL). Additionally, each DLL is restricted to only one routine. Thus, in order to provide a functional API, the definition of several DLLs that access a singleton instance of the native middleware was needed. Also, the inability to define callback rou-




Figure 3: The left picture depicts the sensor middleware and Matlab interface architecture. The right picture shows the Matlab access to read channel MEX routine.

times to push acquired data to the Matlab environment forced us to specify an intermediate data buffering layer. As a result, the process of subscription of a certain service through this interface triggers the creation of a ring buffer for temporary service data storage. In order to access the subscribed service data a polling approach is necessary. To this end, a read channel MEX routine must be called from within Matlab (Fig. 3) that retrieves the data temporarily stored.

Further functionalities are made available at the Matlab API level, such as the connection/disconnection as well as the detection of available devices, service subscription management, namely subscribe/unsubscribe routines as well as the listing of subscribed and non-subscribed services. Also the ability to retrieve status information on a given subscription such as the service signal frequency or the number of available temporary stored samples are provided. Service log management routines are also present.

3 DATA MANAGEMENT

3.1 Persistent Storage

In order to keep record of past acquisitions, a persistent storage must be present along with the application. The database structure can be very simple if one takes into account some assumptions to be true. The assumptions are related to patient information like age, weight, body mass index (BMI), etc., that even though they are dynamic and change over time, their variation is not going to be relevant. Besides the patient information, the database also saves data from collected acquisitions, such as the signals acquired and their processing results, and from any pathologies

that may be found in the acquisition.

The database engine system chosen to provide persistent storage was SQLite (SQLite, 2010). The motivations behind this decision are that this engine is open-source, small and quite reliable. Moreover, it is only a single file which makes the backups operation easier and the ability to accept binary large objects (BLOBs) was mandatory.

This engine is, nonetheless, implemented in C++ and an interface between SQLite and Matlab was necessary. Rather than building our own interface, we have decided to use the one found in (mksqlite, 2010). This interface is able to perform almost every SQL command but lacks the ability to transfer BLOBs. Concerning this issue, we have adapted the original interface to accept BLOBs.

3.2 Import/Export

Another property of the data management module is the faculty to import and export patient's records. The format used to carry the information is XML (eXtensible Markup Language) and Matlab already provides means to parse such documents. A patient record consists of not only personal information but also all the acquisitions' data collected and all the pathologies information found in every single acquisition. Figure 4 provides a small extract of the respective XSD. Among other characteristics, a XSD file gives a basic overview of the final XML structure.

This feature enables the transfer between computers with the application so that, for instance, relevant cases can be shared for training.

```

<?xml version="1.0" encoding="UTF-8"?>
<xs:schema
xmlns:xs="http://www.w3.org/2001/XMLSchema"
elementFormDefault="qualified">
(...)
<xs:element name="acquisition">
  <xs:complexType>
    <xs:sequence>
      <xs:element ref="date"/>
      <xs:element ref="location"/>
      <xs:element ref="position"/>
      <xs:element ref="notes"/>
      <xs:element ref="prim_sound"/>
      <xs:element ref="ref_sound"/>
      <xs:element ref="ecg"/>
      <xs:element ref="noise"/>
      (...)
    </xs:sequence>
    <xs:attribute name="id" use="required"
type="xs:integer"/>
  </xs:complexType>
</xs:element>
(...)
<xs:element name="prim_sound">
  <xs:complexType>
    <xs:sequence>
      <xs:element ref="fs"/>
      <xs:element ref="data"/>
    </xs:sequence>
  </xs:complexType>
</xs:element>
(...)
<xs:element name="data">
  <xs:complexType>
    <xs:sequence>
      <xs:element
maxOccurs="unbounded" ref="sample"/>
    </xs:sequence>
  </xs:complexType>
</xs:element>
(...)
</xs:schema>

```

Figure 4: XSD schema extract for acquisition data.

4 USER INTERFACE

The Graphical User Interface (GUI) works as the "middleman" between the user and the rest of the application modules. The GUI allows the user to:

- manage patient data (new patient record; save acquisition, etc);
- make new acquisitions from a digital stethoscope or load sound clips from a database or from another application;
- configure processing operations to be performed;
- visualise the acquired signals;
- examine the results returned from the processing module;

4.1 Signal Representation

Rather than viewing raw, unprocessed plots of PCG signals which are complex to interpret, a simpler graphical representation is favoured in order to represent the PCG. This representation (see figure 5) seems to be acknowledged not only in the eHealth field (Jiang and Choi, 2006; Tovar-Corona and Torrey, 1997; Reed et al., 2009) but also in traditional medical field (Shaver et al., 1990; Karnath and Thornton, 2002).




Figure 5: Diagram representation of the heart sound signal.

With this representation, heart sound components found during segmentation are graphically displayed with their estimated start and end times, reducing the burden that a physician would have to cope when visualising a PCG. The graphical representation scheme adopted in this work is the one suggested by (Reed et al., 2009; Shaver et al., 1990).

4.2 Acquisition Window

Using the aforementioned signal acquisition toolbox, and along with a digital stethoscope, one is able to acquire new signals. Beforehand, the user may provide information such as the patient position and where the auscultation will take place (see figure 6), by choosing from a preset of possible options. These are relevant since different heart sound components can be best heard in specific positions/locations.

As soon as the acquisition ends, the signals acquired are transferred to the signal processing toolbox. By default there is an already defined analysis workflow. Nevertheless, the user can select and parameterize a subset of processing algorithms in advance.

4.3 Report Window

After the cardiac signal is processed, the physician has several ways to examine the acquisition and its resulting computation values.

In the main window (figure 2) the physician (i) can auscultate the heart sound signal, (ii) examine either the raw cardiac signal or its respective diagram representation and (iii) find the processing values annotated beat by beat.




Figure 6: Real-Time visual representation of heart sound signal acquisition, is depicted on the left figure. The right figure is the acquisition window. The user can select the (i) auscultation location, (ii) patient position, (iii) channels to acquire and (iv) the processing workflow.

Moreover, there is also a report (figure 7) which summarises the complete acquisition in a single window. The report displays (i) patient information, (ii) acquisition information and its processing results with their mean values and respective standard deviation, and (iii) the plot of the cardiac signal with its annotations, providing an overview of the exam.

5 Conclusion and Future Work

Fewer and fewer physicians master the art of cardiac auscultation. In this paper we propose a Matlab tool to support physicians in performing auscultation. This application enables real time signal acquisition using off-the-shelf sensors and performs several automatic annotation functions of heart sounds, such as noise contamination detection, segmentation into S1, S2 and S3, S2-split detection, murmur detection and classification, systolic time intervals measurement, contractility and stroke volume.

Although other clinical applications exist in the literature, they lack some important features like more complex signal processing modules that help supporting the physicians decision. The proposed user interface enables the physician to evaluate the auscultation exam in several options: the physician can view the raw signal as well as its beat-by-beat annotated version. The later applies the graphical representation suggested by (Reed et al., 2009). Finally, a report functionality has been incorporated that presents an overall overview of the exam using numerical average values.

As future work, we expect to deploy the application in a clinical environment for a pilot study in order to evaluate its effectiveness in decision support in daily clinical practice as well as a learning tool to improve auscultation proficiency.

ACKNOWLEDGEMENTS

This work was supported in part by SoundForLife (PTDC/EIA/68620/2006; FCOMP-01-0124-FEDER-007243) financed by the Portuguese Foundation for Science and Technology.

REFERENCES

- Allender, S., Scarborough, P., Peto, V., and Rayner, M. (2008). European cardiovascular disease statistics. *British Heart Foundation Health Promotion*.
- Brito, M., Vale, L., Carvalho, P., and Henriques, J. (2010). A sensor middleware for integration of heterogeneous medical devices. In *Engineering in Medicine and Biology Society, 2010. EMBC 2010. Annual International Conference of the IEEE*.
- Carvalho, P., Paiva, R. P., Kumar, D., Ramos, J. P., Santos, S., and Henriques, J. (2011). A framework for acoustic cardiac signal analysis. *BioStec: BioSignals 2011 (accepted)*.
- Jiang, Z. and Choi, S. (2006). A cardiac sound characteristic waveform method for in-home heart disorder monitoring with electric stethoscope. *Expert Systems with Applications*, 31(2):286 – 298.
- Karnath, B. and Thornton, W. (2002). Auscultation of the heart. *Hospital Physician*, pages 39–43.
- Kudriavtsev, V., Polyshchuk, V., and Roy, D. (2007). Heart energy signature spectrogram for cardiovascular diagnosis. *BioMedical Engineering OnLine*, 6(1):16.
- Lam, M. Z. C., Lee, T. J., Boey, P. Y., Ng, W. F., Hey, H. W., Ho, K. Y., and Cheong, P. Y. (2005). Factors influencing cardiac auscultation proficiency in physician trainees. *Singapore Med J*, 46(1):11–4.
- Mangione, S., Nieman, L. Z., Gracely, E., and Kaye, D. (1993). The Teaching and Practice of Cardiac Auscultation during Internal Medicine and Cardiology Training: A Nationwide Survey. *Annals of Internal Medicine*, 119(1):47–54.
- mksqlite (2010). http://mksqlite.berlios.de/mksqlite_eng.html/.




Figure 7: Report Window. Summarises an acquisition in a single window, giving both patient and auscultation information and also, the computation results.

Rajan, S., Doraiswami, R., Stevenson, R., and Watrous, R. (1998). Wavelet based bank of correlators approach for phonocardiogram signal classification. In *Time-Frequency and Time-Scale Analysis, 1998. Proceedings of the IEEE-SP International Symposium on*, pages 77–80.

Reed, N., Nie, Y., and Mahnke, C. (2009). A portable graphical representation tool for phonocardiograms. In *Engineering in Medicine and Biology Society, 2009. EMBC 2009. Annual International Conference of the IEEE*, pages 3111–3114.

Shaver, J. A., Leonard, J. J., and Leon, D. F. (1990). *Auscultation of the Heart - Examination of the Heart*, volume 4. American Heart Association.

Shub, C. (2003). Echocardiography or auscultation? how to evaluate systolic murmurs. *Can Fam Physician*, 49:163–7.

SQLite (2010). <http://www.sqlite.org/>.

Tovar-Corona, B. and Torry, J. (1997). Graphical representation of heart sounds and murmurs. In *Computers in Cardiology 1997*, pages 101–104.

Vukanovic-Criley, J. M., Criley, S., Warde, C. M., Boker, J. R., Guevara-Matheus, L., Churchill, W. H., Nelson, W. P., and Criley, J. M. (2006). Competency in cardiac examination skills in medical students, trainees, physicians, and faculty: a multicenter study. *Arch Intern Med*, 166(6):610–6.

Watrous, R. L. (2006). Computer-aided auscultation of the heart: from anatomy and physiology to diagnostic decision support. *Conf Proc IEEE Eng Med Biol Soc*, 1:140–3.